
**Surface active agents — Determination
of 1,4-dioxan residues in surfactants
obtained from epoxyethane by gas
chromatography**

*Agents de surface — Détermination de la teneur en résidus
de 1,4-dioxane dans les agents tensioactifs obtenus à partir
d'époxyéthane par chromatographie en phase gazeuse*





COPYRIGHT PROTECTED DOCUMENT

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Principle	1
4 Reagents	1
4.1 1,4-Dioxane, purity 99,5 % minimum	1
4.2 4-Methyl-1,3-dioxane, purity 98 % minimum	1
4.3 N, N-dimethylformamide (DMF)	1
4.4 Carrier gas: nitrogen or helium, purity 99,99 % minimum	2
5 Apparatus	2
5.1 Analytical balance, 0,1 mg readability	2
5.2 Gas chromatography system, attached with temperature programmed device and flame ionization detector (FID)	2
5.3 Head space sampling unit	2
5.4 20 ml crimped top glass vials, sealing pliers and sealing clamp suitable for head space unit	2
5.5 Septum and crimped tops, lined with aluminium or PTFE, for closing the vials	2
5.6 Sealing pliers, for the vials	2
5.7 Capillary column, capable of the separation characteristics shown in Figure 2	2
5.8 Volumetric flasks, 50 ml and 100 ml capacity	2
5.9 Gas tight syringes or glass pipettes, 1 ml $\pm$ 0,01 ml capacity	2
6 Preparation of calibration solutions	2
6.1 Calibration solution A	2
6.2 Calibration solution B	3
7 Procedure	3
7.1 Preparation of the test sample	3
7.2 Standard addition procedure	3
7.2.1 Preparation of test sample solutions	3
7.2.2 Determination	4
7.2.3 Expression of results	4
7.3 Relative response factor procedure	4
7.3.1 Determination of relative response factor	4
7.3.2 Preparation of test sample solutions	5
7.3.3 Determination	5
7.3.4 Expression of results	5
8 Chromatographic conditions	5
9 Evaluation	6
9.1 General	6
9.2 Identification	7
9.3 Calculation	7
10 Precision	7
10.1 Repeatability limit	7
10.2 Reproducibility limit	7
11 Test report	7
Annex A (informative) Statistical data from the ring carried out	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 91, *Surface active agents*.

Surface active agents — Determination of 1,4-dioxan residues in surfactants obtained from epoxyethane by gas chromatography

1 Scope

This International Standard provides a method for the determination of 1,4-dioxane residues in surfactants which are synthesized from epoxyethane, such as alkyl ether sulfates and alcohol ethoxylates.

The method is applicable for samples containing 1,4-dioxane more than 5 mg/kg. For samples containing 1,4-dioxane more than 100 mg/kg, the sample solutions should be diluted to appropriate concentration.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 607, *Surface active agents and detergents — Methods of sample division*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

3 Principle

The sample is weighed into a head space vial and a calibration solution is added. The sealed vial is placed in a head space sampling instrument and allowed to reach thermal equilibrium. A portion of the vapour phase is then analysed by temperature programmed GC with flame ionization detection. Either an internal standard or standard addition method of quantification are employed.

4 Reagents

During the analysis, use only reagents of recognized analytical grade and the water used shall conform to grade 3 in accordance with ISO 3696.

4.1 1,4-Dioxane, purity 99,5 % minimum.

4.2 4-Methyl-1,3-dioxane, purity 98 % minimum.

4.3 N, N-dimethylformamide (DMF).

DMF may be substituted with other solvent. Applicability of the solvent should be evaluated by appropriate validation. Possible substitutes are including, but not limited to following:

- N, N-dimethylacetamide(DMA);
- Dimethyl sulfoxide (DMSO);
- Diethyl formamide (DEF);