BRITISH STANDARD

BS EN
50055:1999

Electrical apparatus for the detection and measurement of combustible gases — Performance requirements for Group I apparatus indicating up to 5 % (*V/V*) methane in air

The European Standard EN 50055:1998 has the status of a British Standard

ICS 13.320

NO COPYING WITHOUT BSI PERMISSION EXCEPT AS PERMITTED BY COPYRIGHT LAW

National foreword

This British Standard is the English language version of EN 50055:1998. It supersedes BS EN 50055:1991 which is withdrawn.

The UK participation in its preparation was entrusted by Technical Committee GEL/31, Electrical apparatus for explosive atmospheres, to Subcommittee GEL/31/19, Gas detectors, which has the responsibility to:

— aid enquirers to understand the text;

— present to the responsible European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;

— monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled "International Standards Correspondence Index", or by using the "Find" facility of the BSI Standards Electronic Catalogue.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, the EN title page, pages 2 to 5 and a back cover.

This British Standard, having been prepared under the direction of the Electrotechnical Sector Committee, was published under the authority of the Standards Committee and comes into effect on 15 May 1999

© BSI 05-1999

ISBN 0 580 30671 2

Amendments issued since publication

Amd. No.	Date	Text affected

EUROPEAN STANDARD

NORME EUROPÉENNE

EUROPÄISCHE NORM

EN 50055

July 1998

ICS 13.220.60; 19.080

Descriptors: Electrical apparatus, explosive atmosphere, mine susceptible to firedamp, detector, measuring apparatus, flammable gas, combustible gas, characteristic

English version

Electrical apparatus for the detection and measurement of combustible gases — Performance requirements for Group I apparatus indicating up to 5 % (*V/V*) methane in air

Appareils électriques de détection et de mesure des gaz combustibles — Régles de performances des appareils du Groupe I pouvant indiquer jusqu'à 5 % (*V/V*) de méthane dans l'air

Elektrische Geräte für das Aufspüren und die Messung brennbarer Gase — Anforderungen an das Betriebsverhalten von Geräten der Gruppe I mit einem Meßbereich bis zu 5 % (*V/V*) Methan in Luft

This European Standard was approved by CENELEC on 12 September 1996.

CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B-1050 Brussels

© 1998 CENELEC All rights of exploitation in any form and by any means reserved worldwide for CENELEC members.

Ref. No. EN 50055:1998 E

Foreword

This second edition of the European Standard was prepared by SC 31-9, Electrical apparatus for the detection and measurement of combustible gases to be used in industrial and commercial potentially explosive atmospheres, of Technical Committee CENELEC TC 31, Electrical apparatus for explosive atmosphere, on the basis of EN 50055:1991, its amendment A1:1995 and a second amendment.

This second amendment was approved by CENELEC on 1996-12-09 for incorporation into a new edition of EN 50055.

This European Standard replaces EN 50055:1991 +A1:1995.

The following dates were fixed:

— latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 1999-01-01

— latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 1999-01-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and covers essential requirements of EC Directive 94/9/EC.

Contents

		Page
Foreword		2
1	Scope	3
2	Normative references	3
3	Definitions	3
4	General requirements	3
5	Performance requirements	3
5.1	General	3
5.2	Unpowered storage	3
5.3	Calibration curve (not applicable to alarm-only apparatus)	3
5.4	Drift (continuous duty apparatus)	3
5.5	Drift (spot reading apparatus)	3
5.6	Alarm	3
5.7	Temperature	3
5.8	Pressure	3
5.9	Humidity	3
5.10	Air speed	3
5.11	Pumping rate	3
5.12	Orientation	3
5.13	Vibration (applicable only to machine-mounted apparatus)	4
5.14	Drop test (applicable to portable apparatus and remote sensors)	4
5.15	Warm-up time (not applicable to spot-reading apparatus)	4
5.15.1	Fixed and transportable apparatus	4
5.15.2	Continuous duty transportable apparatus	4
5.16	Time of response (not applicable to spot-reading apparatus)	4
5.17	Minimum time of operation (spot reading apparatus)	4
5.18	High gas concentrations above the measuring range	4
5.18.1	Non-ambiguity test	4
5.18.2	Residual effect	4
5.19	Battery capacity	4
5.19.1	Battery-powered portable continuous duty apparatus	4
5.19.2	Portable spot-reading apparatus	4
5.20	Power supply variations	4
5.20.1	A.C. powered apparatus	4
5.20.2	External d.c. powered apparatus	4
5.20.3	Battery powered apparatus	4
5.21	Power supply interruptions, voltage transients and step changes of voltage	5
5.22	Addition of sampling probe	5
5.23	Dust	5
5.24	Poisons and other gases	5
5.25	Electromagnetic immunity	5
6	Field calibration kit	5
7	Information for use	5

© BSI 05-1999

1 Scope

1.1 This European Standard specifies performance requirements for Group I (as defined in European Standard EN 50054) portable, transportable and fixed apparatus for the detection and measurement of methane concentrations in mine air. The apparatus, or parts thereof, are intended for use in mines susceptible to firedamps and shall meet the general requirements and test methods specified in EN 50054.

1.2 This European Standard is restricted to apparatus intended for the detection and measurement of methane volume ratios in air from 0 % (*V/V*) up to but not exceeding 5 % (*V/V*).

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to the European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN 50054 1998, *Electrical apparatus for the detection and measurement of combustible gases — General requirements and test methods.*

3 Definitions

For the purposes of this European Standard, the definitions given in EN 50054 apply.

4 General requirements

The apparatus shall comply with the general requirements specified in EN 50054 and with the performance requirements specified in clause **4** of this European Standard.

Compliance shall be determined in accordance with the appropriate test requirements and methods, including initial calibration, specified in EN 50054.

5 Performance requirements

5.1 General

The normal conditions for test are specified in subclause **5.3** of EN 50054. Compliance shall be determined in accordance with the test methods specified in subclause **5.4** of EN 50054.

5.2 Unpowered storage

After being submitted to the conditions specified in subclause **5.4.2** of EN 50054, the apparatus shall meet the requirements specified in **5.3** to **5.25** of this European Standard.

5.3 Calibration curve (not applicable to alarm-only apparatus)

After initial adjustment with the standard test gas, each of the three indications (after correction using the manufacturer's calibration curve, if necessary) obtained for each of four gas volume ratios distributed over the measuring range shall not differ from these volume ratios by more than ±0,1 % (*V/V*) methane or ±5 % of the indication, whichever is the greater.

5.4 Drift (continuous duty apparatus)

The medium term variation shall not exceed ±0,1 % (*V/V*) methane or ±5 % of the indication, whichever is the greater, in air and in the standard test gas.

In addition, the apparatus shall be run in a methane-air mixture [methane volume ratio 1,0 % (*V/V*) to 1,5 % (*V/V*)] for 5 days, readings being taken daily in clean air and in the standard test gas and the variation of the indication in the standard test gas shall not exceed ±0,1 % (*V/V*) methane.

5.5 Drift (spot reading apparatus)

The variation of the indication shall not exceed ±0,1 % (*V/V*) methane or 5 % of the indication, whichever is the greater, in air and in the standard test gas.

5.6 Alarm

The alarm shall operate during every cycle of the test. If a latching alarm is provided, the manual reset action shall be checked during every cycle.

5.7 Temperature

The variation of the indication from that at 20 °C, over the temperature range −10 °C to +40 °C, shall not exceed ±0,2 % (*V/V*) methane or ±10 % of the indication, whichever is the greater, in air and in the standard test gas.

5.8 Pressure

The variation of the indication from that at 101,3 kPa, over a pressure range of 92 kPa to 115 kPa, shall not exceed ±0,25 % (*V/V*) methane or ±15 % of the indication, whichever is the greater, in air and in the standard test gas.

5.9 Humidity

The variation of the indication over a humidity range of 5 % relative humidity (r.h.) to 90 % r.h., at +40 °C, shall not exceed ±0,2 % (*V/V*) methane or ±10 % of the indication, whichever is the greater.

5.10 Air speed

The variation of the indication shall not exceed ±0,1 % (*V/V*) methane or ±5 % of the indication whichever is the greater.

5.11 Pumping rate

The variation of the indication shall not exceed ±0,1 % (*V/V*) methane or ±5 % of the indication, whichever is the greater.

5.12 Orientation

The variation of the indication shall not exceed +0,1 % (*V/V*) methane or ±5 % of the indication, whichever is the greater.

© BSI 05-1999