BS EN ISO 2719:2016+A1:2021
BS 2000-34:2016+A1:2021



BSI Standards Publication

Determination of flash point — Pensky-Martens closed cup method

bsi.

National foreword

This British Standard is the UK implementation of EN ISO 2719:2016+A1:2021. It is identical to ISO 2719:2016, including amendment 1:2021. It is dual numbered as BS 2000-34:2016+A1:2021 in the UK. It supersedes BS EN ISO 2719:2016 (dual numbered as BS 2000-34:2016), which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by ▌A1〉 〈A1▐.

The UK participation in its preparation was entrusted to Technical Committee PTI/13, Petroleum Testing and Terminology.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

© The British Standards Institution 2021
Published by BSI Standards Limited 2021

ISBN 978 0 539 12850 5

ICS 75.080

Compliance with a British Standard cannot confer immunity from legal obligations.

BS 2000 Series

Energy Institute, under the brand of IP, publishes and sells all Parts of BS 2000, and all BS EN and BS ISO petroleum test methods that would be part of BS 2000, both in its annual publication "IP Standard Test Methods for analysis and testing of petroleum and related products, and British Standard 2000 Parts" and individually.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2016.

Amendments/corrigenda issued since publication

Date	Text affected
30 April 2021	Implementation of ISO amendment 1:2021 to BS EN ISO 2719:2016, still dual numbered as BS 2000-34:2016, with CEN endorsement A1:2021

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 2719:2016+A1

March 2021

ICS 75.080

English Version

Determination of flash point - Pensky-Martens closed cup method (ISO 2719:2016)

Détermination du point d'éclair - Méthode Pensky-Martens en vase clos (ISO 2719:2016)

Bestimmung des Flammpunktes - Verfahren nach Pensky-Martens mit geschlossenem Tiegeln (ISO 2719:2016)

This European Standard was approved by CEN on 10 July 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

© 2021 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members.

Ref. No. EN ISO 2719:2016/A1:2021 E

European foreword

This document (EN ISO 2719:2016) has been prepared by Technical Committee ISO/TC 28 "Petroleum products and lubricants" in collaboration Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin” the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2017, and conflicting national standards shall be withdrawn at the latest by January 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 2719:2002.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 2719:2016 has been approved by CEN as EN ISO 2719:2016 without any modification.

Foreword to amendment A1

This document (EN ISO 2719:2016/A1:2021) has been prepared by Technical Committee ISO/TC 28 "Petroleum and related products, fuels and lubricants from natural or synthetic sources" in collaboration with Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin.” the secretariat of which is held by NEN.

This Amendment to the European Standard EN ISO 2719:2016 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2021, and conflicting national standards shall be withdrawn at the latest by September 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 2719:2016/Amd 1:2021 has been approved by CEN as EN ISO 2719:2016/A1:2021 without any modification.

Contents

Page

Foreword ... **iv**
Introduction ... **v**
1 Scope ... **1**
2 Normative references ... **1**
3 Terms and definitions ... **2**
4 Principle ... **2**
5 Chemicals and materials ... **2**
6 Apparatus ... **2**
7 Apparatus preparation ... **3**
7.1 General ... 3
7.2 Location of the apparatus ... 3
7.3 Cleaning the test cup ... 3
7.4 Apparatus assembly ... 3
7.5 Apparatus verification ... 3
8 Sampling ... **4**
9 Sample handling ... **4**
9.1 Petroleum products ... 4
9.1.1 Sub-sampling ... 4
9.1.2 Samples containing undissolved water ... 4
9.1.3 Samples that are liquid at ambient temperature ... 4
9.1.4 Samples that are very viscous, semi-solid or solid at ambient temperature ... 4
9.2 Paints and varnishes ... 5
10 Procedure ... **5**
10.1 General ... 5
10.2 Procedure A ... 5
10.3 Procedure B ... 6
10.4 Procedure C ... 7
11 Calculation ... **8**
11.1 Conversion of barometric pressure reading ... 8
11.2 Correction of observed flash point to standard atmospheric pressure ... 8
12 Expression of results ... **8**
13 Precision ... **8**
13.1 General ... 8
13.2 Repeatability, ***r*** ... ***8***
13.3 Reproducibility, ***R*** ... ***9***
14 Test report ... **10**
Annex A (informative) **Apparatus verification** ... **11**
Annex B (normative) **Pensky-Martens closed cup test apparatus** ... **14**
Annex C (normative) **Temperature measuring device specification** ... **20**
Bibliography ... **22**

© ISO 2016 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 28, *Petroleum products and related products of synthetic or biological origin*, in conjunction with ISO/TC 35, *Paints and varnishes*.

This fourth edition cancels and replaces the third edition (ISO 2719:2002), which has been technically revised.

The main technical updates include:

a) introduction of procedure C for FAME products;

b) revision of temperature measuring device requirements, allowing alternatives for Hg-containing thermometers;

c) removal of the original Annex D on the adaptor for the low-range thermometer, it being optional when a permanent ferrule is attached to the thermometer;

d) revision of procedures regarding sampling and sample handling;

e) inclusion, for automated apparatus, manufactured after 1 January 2017, of a device to automatically dispense an inert gas or vapour over the test cup in the event of a test cup fire.

© ISO 2016 – All rights reserved

Introduction

Flash point values are used in shipping, storage, handling, and safety regulations, as a classification property to define “flammable” and “combustible” materials. Precise definition of the classes is given in each particular regulation.

A flash point value can indicate the presence of highly volatile material(s) in a relatively non-volatile or non-flammable material and flash point testing can be a preliminary step to other investigations into the composition of unknown materials.

It is not appropriate for flash point determinations to be carried out on potentially unstable, decomposable, or explosive materials, unless it has been previously established that heating the specified quantity of such materials in contact with the metallic components of the flash point apparatus, within the temperature range required for the method, does not induce decomposition, explosion or other adverse effects.

Flash point values are not a constant physical-chemical property of materials tested. They are a function of the apparatus design, the condition of the apparatus used, and the operational procedure carried out. Flash point can therefore be defined only in terms of a standard test method, and no general valid correlation can be guaranteed between results obtained by different test methods or with test apparatus different from that specified.

ISO/TR 29662[6] (CEN/TR 15138[12]) gives useful advice in carrying out flash point tests and interpreting their results.

© ISO 2016 – All rights reserved

INTERNATIONAL STANDARD ISO 2719:2016+A1:2021(E)

Determination of flash point — Pensky-Martens closed cup method

WARNING — The use of this International Standard can involve hazardous materials, operations and equipment. This International Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of users of this International Standard to take appropriate measures to ensure the safety and health of personnel prior to application of this International Standard, and fulfil statutory and regulatory requirements for this purpose.

1 Scope

This International Standard describes three procedures, A, B and C, using the Pensky-Martens closed cup tester, for determining the flash point of combustible liquids, liquids with suspended solids, liquids that tend to form a surface film under the test conditions, biodiesel and other liquids in the temperature range of 40 °C to 370 °C.

CAUTION — For certain mixtures no flash point, as defined, is observed; instead a significant enlargement of the test flame (not halo effect) and a change in colour of the test flame from blue to yellowish-orange can occur. Continued heating can result in significant burning of vapours outside the test cup, and can be a potential fire hazard.

NOTE 1 Although, technically, kerosene with a flash point above 40 °C can be tested using this International Standard, it is standard practice to test kerosene according to ISO 13736.[5] Similarly, lubricating oils are normally tested according to ISO 2592[2].

Procedure A is applicable to distillate fuels (diesel, biodiesel blends, heating oil and turbine fuels), new and in-use lubricating oils, paints and varnishes, and other homogeneous liquids not included in the scope of Procedures B or C.

Procedure B is applicable to residual fuel oils, cutback residua, used lubricating oils, mixtures of liquids with solids, liquids that tend to form a surface film under test conditions or are of such kinematic viscosity that they are not uniformly heated under the stirring and heating conditions of Procedure A.

Procedure C is applicable to fatty acid methyl esters (FAME) as specified in specifications such as EN 14214[11] or ASTM D6751[13].

This International Standard is not applicable to water-borne paints and varnishes.

NOTE 2 Water-borne paints and varnishes can be tested using ISO 3679[3]. Liquids containing traces of highly volatile materials can be tested using ISO 1523[1] or ISO 3679.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1513, *Paints and varnishes — Examination and preparation of test samples*

ISO 3170, *Petroleum liquids — Manual sampling*

ISO 3171, *Petroleum liquids — Automatic pipeline sampling*

ISO 15528, *Paints, varnishes and raw materials for paints and varnishes — Sampling*

© ISO 2016 – All rights reserved